

TallPrep™ Adapter

The TallPrep™ Adapter enables the FastPrep-24™ 5G (116005500) and Classic (116004500) instruments to accommodate 24 x 4.5 mL Lysing Matrix tubes. The TallPrep Adapter consists of 2 parts (Fig 1):

- ▶ **Bottom component (116002540B):**
holding base for 4.5 mL Lysing Matrix tubes
- ▶ **Top component (116002540T):**
holding plate to secure tubes and adapter

⚠ ATTENTION ▶ *It is recommended to use either the Ratchet Nut v3.0, blue (116005510) or Cam Lock (116005588), sold separately, with this adapter. If any other nut not compatible with this adapter is used, there is a high risk of damage that may occur to the instrument.*

INSTRUCTIONS FOR USE

- 1 Place the bottom component of the TallPrep Adapter on the 3-Step Assembly. Rotate and align the 3 ridges underneath bottom component to the 3 indents on the 3-Step Assembly of the FastPrep-24 instrument. Light pressure may be necessary to secure the adapter in place (Fig 2).
- 2 Load the 4.5 mL Lysing Matrix tubes into bottom component. Tubes should face upright with a 60 degree inward tilt (Fig 3). If using less than 24 tubes, distribute the load around evenly.
- 3 Place the top component of the TallPrep adapter onto the tube-loaded bottom component. Align the 4 rectangular notches around the pole of the bottom component to the 4 inserts on the top component.

⚠ WARNING ▶ *Failure to secure the adapter as detailed in step 4 may result in PERMANENT and SEVERE adapter damage.*



▶ **DATA SHEET:** TallPrep™ Adapter

- 4 Secure the adapter with either the Ratchet Nut or Cam Lock on the FastPrep-24 instrument.

4.1: Use with Ratchet Nut system

Turn Ratchet Nut clockwise until a clicking sound is heard. Press down and turn the Ratchet Nut for one additional half rotation (Fig 4). Ensure that the adapter does not move and that the Ratchet Nut is resistant to loosening.

4.2: Use with Cam Lock system (for FastPrep-24 5G only)

⚠ WARNING ▶ Never turn Cam Lock on threaded bolt in the locked position.

4.2.1 Place the Cam Lock in the unlocked position onto the threaded bolt (Fig 5). Rotate the Cam Lock latch clockwise to tighten down. Stop when Cam Lock contacts the top component of the adapter and resistance is felt.

4.2.2 Rotate approximately half turn counterclockwise.

4.2.3 Flip up the lever arm to lock and hold the adapter steady with the other hand (Fig 6).

- 5 Close the dome and secure. Select program or use manual setting for speed, time and cycle. Press run key.

⚠ CAUTION ▶ If the locking system becomes loose during the homogenization process, immediately stop the FastPrep-24 instrument by pushing down the emergency STOP BUTTON. Remove the locking system, clean the thread of the threaded bolt and repeat steps 4 and 5. If this repeatedly occurs, replace the locking system or contact MP Biomedicals technical support.

- 6 If consecutive runs are necessary to completely lyse your sample, manually check for heat build-up to the adapter parts after each run. Before starting next run, hold palm of hand on the adapter top and body. If heat is present, allow enough time to cool down in between runs for the heat to dissipate. Failure to perform this step can cause adapter failure and damage.

Ratchet Nut system:



Cam Lock system (FastPrep-24 5G only):



MP BIOMEDICALS

AMERICAS: 800.854.0530 | biotech@mpbio.com

EUROPE: 00800.7777.9999 | techsupport.eur@mpbio.com

APAC: +65 6775.0008 | custserv.ap@mpbio.com

Learn more at www.mpbio.com

