

# MagBeads FastRNA Kit for FFPE (Ready-to- Use for MPure-32)

Magnetic bead-based Purification for total RNA from FFPE  
samples.

**Size:** 96 preps

**Storage:** 15-25 °C

**Cat. No.:** 117033500

**Content Version:** Feb 2024



# Table of Contents

1. Introduction to MagBeads FastRNA Kit for FFPE .....	3
2. Kit Components and User Supplied Materials .....	4
2.1 MagBeads FastRNA Kit for FFPE Component .....	4
2.2 User Supplied Materials .....	4
3. Storage and Kit Stability .....	4
4. Important Consideration Before Use .....	5
5. Safety Precautions .....	6
6. Protocol .....	7
7. Troubleshooting .....	9
8. Product Use Limitation & Warranty .....	10

## 1. Introduction to MagBeads FastRNA Kit for FFPE

MagBeads FastRNA Kit for FFPE is intended for rapid extraction of RNA from tissue, cells, blood, and other clinical samples using the MPure-32™ aNAP System. RNA can be used directly for RT-PCR, quantitative RT-PCR and so on.

MagBeads FastRNA Kit for FFPE is based on the purification method of high binding magnetic particles. The sample is lysed and digested. RNA is released into the lysate. After addition of magnetic particles and binding solution, RNA will be adsorbed on the surface of magnetic particles, and impurities such as proteins will be removed without adsorption. The adsorbed particles were washed with washing buffer to remove the proteins and impurities, washed with ethanol to remove salts, and finally the RNA was eluted with Elution Buffer.

Visit [www.mpbio.com](http://www.mpbio.com) to explore additional products to support your research.

## 2. Kit Components and User Supplied Materials

### 2.1 MagBeads FastRNA Kit for FFPE Component

MagBeads FastRNA Kit for FFPE (#117033500, 192 Preps)	
Components	Package
96-Well Reagent Plate	6 plates
Proteinase K	48 mg
Protease Dissolve Buffer	3 mL
DNase I	2 x 600 µL
Buffer DPS	100 ml
Buffer FRL	30 ml
Buffer MW1	22 mL
8-strip A (Cover for Magnetic Rod)	12 pieces

### 2.2 User Supplied Materials

- Disposable powder-free gloves.
- Pipettes (adjustable).
- Sterile pipette tips with aerosol barriers (up to 200 µl).
- Vortex mixer.
- Desktop microcentrifuge with rotor for 2 ml reaction tubes (RCF max. 16,000 x g).
- PCR box or Biological cabinet. • Vacuum aspirator with flask for removing supernatant.
- Tube racks.
- 1.5 ml polypropylene sterile tubes.
- Refrigerator for 2-8°C.
- Deep-freezer for ≤ -16°C.
- Waste bin for used tips.
- Permanent pen for labeling
- Thermostatic bath or dry block for tubes with controlled temperature and capable of incubating at 25-100°C.

## 3. Storage and Kit Stability

Proteinase K, DNase I, Magbeads Particles should be stored at 2-8°C upon arrival. However, short-term storage (up to 24 weeks) at room temperature (15-25°C) does not

affect their performance. The remaining kit components can be stored at room temperature (15-25°C) and are stable for at least 18 months under these conditions.

## 4. Important Consideration Before Use

- ❑ Add 2.4 mL Protease Dissolve Buffer into Proteinase K bottle, and store at -20°C after it dissolves.

## 5. Safety Precautions

The user should always pay attention to the following:

- Use sterile pipette tips with aerosol filters and use new tip for every procedure.
- Store extracted positive material (samples, controls and amplicons) away from all other reagents.
- Thaw all components thoroughly at room temperature before starting an assay.
- When thawed, mix the components and centrifuge briefly.
- Use disposable gloves, laboratory coats, protect eyes while samples and reagents handling. Thoroughly wash hands afterwards.
- Do not eat, drink, smoke, apply cosmetics, or handle contact lenses in laboratory work areas.
- Do not use a kit after its expiry date.
- Dispose of all samples and unused reagents in compliance with local authorities requirements.
- Samples should be considered potentially infectious and handled in a biological cabinet in compliance with appropriate biosafety practices.
- Clean and disinfect all sample or reagent spills using a disinfectant such as 0.5% sodium hypochlorite, or other suitable disinfectant.
- Avoid contact with the skin, eyes and mucose membranes. If skin, eyes and mucose membranes contact immediately flush with water, seek medical attention.
- Material Safety Data Sheets (MSDS) are available on request.
- Use of this product should be limited to personnel trained in the techniques of DNA amplification.
- The laboratory process must be one directional; it should begin in the Extraction Area move to the Amplification and Detection Area. Do not return samples, equipment and reagents to the area in which the previous step was performed.

## 6. Protocol

### Sample Preparation

- Using a scalpel, trim off the excess paraffin on the sample block. Cut up to 8 sections (5~20  $\mu\text{m}$ ) and immediately transfer samples into a 2 mL microcentrifuge tube or a 2 mL 96-well sample plate (not provided).

**Note:** Discard the first 2 to 3 sections of sample if it has been exposed to air.

- Add 600  $\mu\text{L}$  Buffer DPS into the sample tube. Vortex the tube for 5s and briefly centrifuge to bring down the sample.
- Incubate at 56 °C for 3 to 5 mins and vortex for 5s to dissolve the paraffin.

**Note:** The Buffer DPS may turn opaque or cloudy. If this occurs, add additional Buffer DPS and repeat Step 2.

- Centrifuge at 14,000 x g for 2 mins and carefully discard the supernatant without disturbing the pellet.
- Add 150  $\mu\text{L}$  Buffer FRL and 20  $\mu\text{L}$  Proteinase K into the sample and vortex. Incubate at 55°C for 15 mins. then incubate at 80°C for 15 mins. Briefly centrifuge the sample and proceed to extraction step.

**Note:** Incubation at 80°C can reverse the nucleic acids modified by formaldehyde. Prolonged incubation time will cause degradation of RNA.

### Automation Extraction with MPure-32

- Transfer 150  $\mu\text{L}$  of sample to well #1 or #7 of the pre-filled reagent.
- Add 10  $\mu\text{L}$  of DNase I into well #3 or #9.
- Place the reagent plate on MPure-32™ aNAP System and run the assay with the program named “FFPE\_RNA” which has the following setting:

Step	Well	Process	Time (s)			Mixing Speed	Temp (C)
			Mix	Wait	Attract		
1	#1/#7	Bind	600	300	120	Medium	RT
2	#2/#8	Wash 1	90	0	60	Medium	RT
3	#3/#9	DNase	60	0	60	Medium	RT
4	#4/#10	Wash 2	60	0	60	Medium	RT
5	#5/#11	Wash 3	60	0	60	Medium	RT
6	#5/#11	Dry	0	300	0	-	RT
7	#6/#12	Elute	300	0	120	Medium	RT
8	#1/#7	Magbeads Release	60	0	0	Medium	RT

9. Transfer the clear purified RNA to a clean 96-well microplate (not provided). The eluent is now ready for downstream applications. Store the purified RNA at  $-80^{\circ}\text{C}$  for an extended storage.

Note: If there are still Magnetic Beads remaining in eluent, please centrifuge at  $14,000 \times g$  for 3-5 mins and transfer the supernatant to a clean 96-well microplate or 1.5 mL microcentrifuge tubes.



## 7. Troubleshooting

This guide may be useful in solving any problems that may arise. For further assistance, please contact our technical support team at [apac-techsupport@mpbio.com](mailto:apac-techsupport@mpbio.com)

Problem	Recommendation
False negatives with extraction product	Degradation of the nucleic acid contained in the sample. Use a new sample, store samples appropriately.
	Loss of nucleic acid deposit. Carefully draw off the wash solution and try not to remove the nucleic acid deposit.
	Degradation of the extracted nucleic acid. Plastic free from DNAses and RNAses should be used. Use a new aliquot of kit's component.
False positives with extraction product	Contamination during sample extraction. One test tube at a time should be opened. Avoid spilling the contents of the test tube, always change tips.
	Contamination of the reagents prepared for the step. Use a new aliquot of a component.
	Contamination of the extraction zone by amplicons. Surfaces and instruments using aqueous detergents should be cleaned, wash lab coats, replace test tubes and tips in use.

## 8. Product Use Limitation & Warranty

The products presented in this instruction manual are for research or manufacturing use only. They are not to be used as drugs or medical devices to diagnose, cure, mitigate, treat, or prevent diseases in humans or animals, either as part of an accepted course of therapy or in experimental clinical investigation. These products are not to be used as food, food additives or general household items. Purchase of MP Biomedicals products does not grant rights to reproduce, modify, or repackage the products or any derivative thereof to third parties. MP Biomedicals makes no warranty of any kind, expressed or implied, including merchantability or fitness for any particular purpose, except that the products sold will meet our specifications at the time of delivery.

Buyer's exclusive remedy and the sole liability of MP Biomedicals hereunder shall be limited to, at our discretion, no replacement or compensation, product credits, refund of the purchase price of, or the replacement of materials that do not meet our specification. By acceptance of the product, Buyer indemnifies and holds MP Biomedicals harmless against, and assumes all liability for, the consequence of its use or misuse by the Buyer, its employees, or others, including, but not limited to, the cost of handling. Said refund or replacement is conditioned on Buyer notifying within thirty (30) days of receipt of product. Failure of Buyer to give said notice within thirty (30) days shall constitute a waiver by the Buyer of all claims hereunder with respect to said material(s).

**Australia**

Tel: +61 2.8824.2100  
Tel: +61 1800.249.998  
Email: [custserv.au@mpbio.com](mailto:custserv.au@mpbio.com)

**Austria & Germany**

Tel: 0800.426.67.337  
Tel: 00800.7777.9999  
Email: [custserv.de@mpbio.com](mailto:custserv.de@mpbio.com)

**Belgium**

Tel: 00800.7777.9999  
Email: [custserv.be@mpbio.com](mailto:custserv.be@mpbio.com)

**Canada**

Tel: +1 800.854.0530  
Email: [custserv.ca@mpbio.com](mailto:custserv.ca@mpbio.com)

**China**

Tel: +86 400.150.0680  
Email: [custserv.cn@mpbio.com](mailto:custserv.cn@mpbio.com)

**Europe**

Tel: +33 3.88.67.54.25  
Tel: +33 00800.7777.9999  
Email: [custserv.eur@mpbio.com](mailto:custserv.eur@mpbio.com)

**France**

Tel: +33 3.88.67.54.25  
Email: [custserv.fr@mpbio.com](mailto:custserv.fr@mpbio.com)

**India**

Tel: +91 22.27636921/22/25  
Email: [custserv.in@mpbio.com](mailto:custserv.in@mpbio.com)

**Italy**

Tel: 00800.7777.9999  
Email: [custserv.it@mpbio.com](mailto:custserv.it@mpbio.com)

**Japan**

Tel: +81 3.6667.0730  
Email: [custserv.jp@mpbio.com](mailto:custserv.jp@mpbio.com)

**Latin America**

Tel: +1 800.854.0530  
Tel: +1 440.337.1200  
Email: [custserv.la@mpbio.com](mailto:custserv.la@mpbio.com)

**New Zealand**

Tel: +64 9.912.2460  
Email: [custserv.nz@mpbio.com](mailto:custserv.nz@mpbio.com)

**North America**

Tel: +1 800.854.0530  
Tel: +1 440.337.1200  
Email: [custserv.na@mpbio.com](mailto:custserv.na@mpbio.com)

**Poland**

Tel: 00800.7777.9999  
Email: [custserv.po@mpbio.com](mailto:custserv.po@mpbio.com)

**Russia**

Tel: +7 495 604.13.44  
Email: [custserv.rs@mpbio.com](mailto:custserv.rs@mpbio.com)

**Serbia**

Tel: +381 11.242.1972  
Email: [custserv.se@mpbio.com](mailto:custserv.se@mpbio.com)

**Singapore/ APAC**

Tel: +65 6775.0008  
Tel: +65 6394.7675  
Email: [custserv.ap@mpbio.com](mailto:custserv.ap@mpbio.com)

**South Korea**

Tel: +82 2.425.5991  
Email: [custserv.kr@mpbio.com](mailto:custserv.kr@mpbio.com)

**Switzerland**

Tel: 00800.7777.9999  
Email: [custserv.ch@mpbio.com](mailto:custserv.ch@mpbio.com)

**The Netherlands**

Tel: 00800.7777.9999  
Email: [custserv.nl@mpbio.com](mailto:custserv.nl@mpbio.com)

**United Kingdom**

Tel: 0800.282.474  
Email: [custserv.uk@mpbio.com](mailto:custserv.uk@mpbio.com)

[www.mpbio.com](http://www.mpbio.com)

